

## **Your Resuscitation trolley and checklist.**

### **PLEASE READ CAREFULLY.**

This system has been designed to be easier and quicker to check, with the aim of improving Resuscitation trolley compliance and patient safety. Stock has been reduced to only necessary equipment, with a pictorial reference for checking in this file.

The Trolley Top and seal documentation are to be completed **DAILY**.  
Full trolley check and re-seal is to be performed **WEEKLY**.

The Airway and Circulation drawers have been replaced with wrapped trays. **These trays are only to be opened for use in an emergency.** As long as they remain fully wrapped and in date you can consider the drawer **CHECKED**.

The **Airway** tray is wrapped in **blue**



The **Circulation** tray is wrapped in **red**

A spare stock of trays is kept in the Resuscitation Store on **Winter ward**. These can be exchanged if your tray is out of date or has been opened. Return the tray, with its un-used contents, to the store room. We will recycle any unused items.

### **Do not leave any sharps in these trays.**

Should you have any further queries, please contact the Resuscitation Service on 8562, alternatively email [paht.resus@nhs.net](mailto:paht.resus@nhs.net)